**QUEEN'S
UNIVERSITY
BELFAST**

Determination of trace elements and macronutrients in agricultural soils using energy dispersive X-ray fluorescence as a rapid and precise analytical technique

Croffie, M. E. T., Williams, P., Daly, K., Fenton, O., Fenelon, A., & Metzger, K. (2020). *Determination of trace elements and macronutrients in agricultural soils using energy dispersive X-ray fluorescence as a rapid and precise analytical technique*. Abstract from EGU - General Assembly 2020, Vienna, Austria.

Document Version:
Peer reviewed version

Queen's University Belfast - Research Portal:
[Link to publication record in Queen's University Belfast Research Portal](#)

Publisher rights

Copyright 2020 the authors.

This is an open access article published under a Creative Commons Attribution License (<https://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution and reproduction in any medium, provided the author and source are cited.

General rights

Copyright for the publications made accessible via the Queen's University Belfast Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

The Research Portal is Queen's institutional repository that provides access to Queen's research output. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact openaccess@qub.ac.uk.

Determination of trace elements and macronutrients in agricultural soils using energy dispersive X-ray fluorescence as a rapid and precise analytical technique

Maame Ekua Tawiah Croffie ^{1,2}, Paul N. Williams ¹, Owen Fenton ², Anna Fenelon ², Konrad Metzger ^{2,3}, Karen Daly ²

¹ Institute for Global Food Security, School of Biological Sciences, Queen's University Belfast, Northern Ireland, BT9 5DL, United Kingdom

² Environment, Soils and Land Use Department, Teagasc, Johnstown Castle Research Centre, Wexford, Y35 TC97, Ireland

³ International Network for Environment and Health, School of Geography and Archaeology, Department of Geography, Arts and Science Building, NUI Galway, H91 TK33, Ireland

Abstract

X-ray fluorescence spectrometry (XRF) is a rapid and inexpensive method for soil analysis. Although, not as precise and accurate as mineral acid digestion-Inductively Coupled Plasma spectrometry for soil testing at present, the XRF method has the potential to be optimised. The objective of this study was to reduce moisture effects, particle size effects and spectra interferences on trace element and macronutrient analyses by improving soil sample preparation and XRF calibration, using bench top Energy-dispersive X-ray Fluorescence Spectrometry (EDXRF). The soil particle size effects study involved samples prepared as pressed powders, pressed pellets, and pressed pellets with a wax binder. After which, the recoveries of Al, Cr, Ni, Mn, Pb, Ca, Fe, K, Mg, P, S and Zn were evaluated as a measure of accuracy and precision. The XRF was calibrated with the fundamental parameters (FP) and matching library (FPML) methods to reduce spectral interferences and validated with certified reference materials. In addition, both XRF methods (FP and FPML) were compared with *aqua regia* digestion (acid digestion) -Inductively Coupled Plasma-Optic Emission Spectrometry (ICP-OES) using concordance correlation coefficient (CCC), whereby a value of 1 indicated good agreement between methods. There were significant differences ($p < 0.05$) between the sample preparation methods and the pressed pellet with wax binder had the best accuracy and precision for all the elements. In addition, for the calibration study, the FPML gave better recoveries of Ni, Ca, Mg, S, P, Cr compared to the FP, however, for Fe, Zn, K and Mn the FP had better recoveries than the FPML. Furthermore, there was good agreement ($CCC > 0.80$) between both XRF methods and ICP-OES for all elements except Al, P, Cu, K, S and Cr ($CCC < 0.60$). *Aqua regia* digestion underestimated the total concentration of Al and K, thus, the XRF had better accuracy for predicting these elements. However, for Cr, S, Cu and P, the XRF overestimated the concentration of these elements in soil. This study has shown that the XRF is as accurate as the ICP-spectrometry for most elements when properly calibrated with the advantage of having a shorter sample preparation. Thus, environmental analysts can monitor soils frequently using the XRF without losing analytical results especially with Al and K and also incurring substantial costs in analyses as with traditional methods.

Keywords: soil; elemental analysis; X-Ray Fluorescence; spectrometry; calibration